

CALIBRATION CERTIFICATION

We certify that the calibration accuracies listed below are obtained on equipment, and with methods, that can be traced directly to the US National Institute of Standards and Technology.

FLOWRATE READOUT ACCURACY: \pm 1% Full Scale

PRESSURE TESTED AT:

METER SERIAL NUMBER:

MODEL NUMBER:

The calibration listed above was performed under the following conditions:

0 to 600 SCFM SATURATED LANDFILL GAS
32 to 200° F
0 to 15 PSIG
4" SDR 17 HDPE LINE
3.939" LINE ID

Standard Conditions: 60° F @ 1 ATM

Signature _____

Date

THERMAL INSTRUMENT COMPANY, INC.

217 Sterner Mill Road, Trevoese, PA 19053

Phone: 215-355-8400 Fax: 215-355-1789

Web: www.thermalinstrument.com